

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
 Data File : VX001066.D
 Acq On : 23 Apr 2018 18:27
 Operator : JC/MD
 Sample : J2464-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0445-180147

Quant Time: Apr 24 08:07:18 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	206286	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	284629	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	257187	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	152664	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	134504	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
35) Dibromofluoromethane	5.51	113	103388	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
50) Toluene-d8	8.72	98	399259	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	11.14	95	144045	42.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.54%	
Target Compounds						
8) Diethyl Ether	2.19	74	2116	1.54	ug/l	99
27) cis-1,2-Dichloroethene	4.61	96	15195	6.10	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
Data File : VX001066.D
Acq On : 23 Apr 2018 18:27
Operator : JC/MD
Sample : J2464-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0445-180147

Quant Time: Apr 24 08:07:18 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

