

Data Path : W:\HPCHEM1\MSVOA X\Data\VX042318\
 Data File : VX001048.D
 Acq On : 23 Apr 2018 10:21
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 24 07:11:59 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.620	0.666	-7.4	102	0.00
3 P	Chloromethane	0.577	0.546	5.4	96	0.00
4 C	Vinyl Chloride	0.598	0.576	3.7#	99	0.00
5 T	Bromomethane	0.370	0.359	3.0	97	0.00
6 T	Chloroethane	0.356	0.342	3.9	101	0.00
7 T	Trichlorofluoromethane	0.899	0.905	-0.7	104	0.00
8 T	Diethyl Ether	0.332	0.313	5.7	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.527	0.528	-0.2	106	0.00
10 T	Methyl Iodide	0.437	0.452	-3.4	101	0.00
11 T	Tert butyl alcohol	0.096	0.085	11.5	98	-0.01
12 CM	1,1-Dichloroethene	0.487	0.471	3.3#	101	0.00
13 T	Acrolein	0.031	0.042	-35.5#	159#	0.00
14 T	Allyl chloride	0.882	0.863	2.2	102	0.00
15 T	Acrylonitrile	0.259	0.226	12.7	95	0.00
16 T	Acetone	0.239	0.247	-3.3	114	0.00
17 T	Carbon Disulfide	1.388	1.332	4.0	102	0.00
18 T	Methyl Acetate	0.722	0.710	1.7	98	0.00
19 T	Methyl tert-butyl Ether	1.638	1.590	2.9	103	0.00
20 T	Methylene Chloride	0.556	0.500	10.1	101	0.00
21 T	trans-1,2-Dichloroethene	0.542	0.514	5.2	103	0.00
22 T	Diisopropyl ether	1.707	1.656	3.0	100	0.00
23 T	Vinyl Acetate	1.413	1.326	6.2	98	0.00
24 P	1,1-Dichloroethane	0.957	0.907	5.2	101	0.00
25 T	2-Butanone	0.373	0.345	7.5	101	0.00
26 T	2,2-Dichloropropane	0.823	0.867	-5.3	109	0.00
27 T	cis-1,2-Dichloroethene	0.604	0.582	3.6	102	0.00
28 T	Bromochloromethane	0.368	0.377	-2.4	106	-0.01
29 T	Tetrahydrofuran	0.235	0.206	12.3	94	0.00
30 C	Chloroform	0.991	0.974	1.7#	102	0.00
31 T	Cyclohexane	0.894	0.872	2.5	105	0.00
32 T	1,1,1-Trichloroethane	0.889	0.905	-1.8	104	0.00
33 S	1,2-Dichloroethane-d4	0.634	0.568	10.4	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.381	0.358	6.0	94	0.00
36 T	1,1-Dichloropropene	0.588	0.580	1.4	103	0.00
37 T	Ethyl Acetate	0.539	0.499	7.4	96	0.00
38 T	Carbon Tetrachloride	0.595	0.635	-6.7	107	0.00
39 T	Methylcyclohexane	0.684	0.702	-2.6	105	0.00
40 TM	Benzene	1.634	1.626	0.5	102	0.00
41 T	Methacrylonitrile	0.337	0.301	10.7	96	0.00
42 TM	1,2-Dichloroethane	0.627	0.619	1.3	102	0.00
43 T	Isopropyl Acetate	0.925	0.893	3.5	99	0.00
44 TM	Trichloroethene	0.499	0.496	0.6	103	0.00
45 C	1,2-Dichloropropane	0.411	0.416	-1.2#	101	0.00

Data Path : W:\HPCHEM1\MSVOA X\Data\VX042318\
 Data File : VX001048.D
 Acq On : 23 Apr 2018 10:21
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 24 07:11:59 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.287	0.281	2.1	101	0.00
47 T	Bromodichloromethane	0.532	0.574	-7.9	106	0.00
48 T	Methyl methacrylate	0.498	0.472	5.2	97	0.00
49 T	1,4-Dioxane	0.007	0.007	0.0	106	0.00
50 S	Toluene-d8	1.410	1.306	7.4	94	0.00
51 T	4-Methyl-2-Pentanone	0.561	0.528	5.9	96	0.00
52 CM	Toluene	1.066	1.080	-1.3#	103	0.00
53 T	t-1,3-Dichloropropene	0.636	0.697	-9.6	106	0.00
54 T	cis-1,3-Dichloropropene	0.600	0.668	-11.3	106	0.00
55 T	1,1,2-Trichloroethane	0.424	0.422	0.5	102	0.00
56 T	Ethyl methacrylate	0.622	0.625	-0.5	100	0.00
57 T	1,3-Dichloropropane	0.702	0.691	1.6	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.304	0.319	-4.9	104	0.00
59 T	2-Hexanone	0.430	0.414	3.7	99	0.00
60 T	Dibromochloromethane	0.445	0.496	-11.5	106	0.00
61 T	1,2-Dibromoethane	0.447	0.448	-0.2	101	0.00
62 S	4-Bromofluorobenzene	0.592	0.557	5.9	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.531	0.530	0.2	104	0.00
65 PM	Chlorobenzene	1.331	1.328	0.2	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.457	0.490	-7.2	104	0.00
67 C	Ethyl Benzene	2.211	2.242	-1.4#	103	0.00
68 T	m/p-Xylenes	0.871	0.892	-2.4	104	0.00
69 T	o-Xylene	0.840	0.851	-1.3	103	0.00
70 T	Styrene	1.384	1.435	-3.7	103	0.00
71 P	Bromoform	0.375	0.422	-12.5	106	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.269	3.283	-0.4	105	0.00
74 T	N-amyl acetate	1.398	1.311	6.2	99	0.00
75 P	1,1,2,2-Tetrachloroethane	0.925	0.887	4.1	99	0.00
76 T	1,2,3-Trichloropropane	0.867	0.853	1.6	102	0.00
77 T	Bromobenzene	0.949	0.928	2.2	102	0.00
78 T	n-propylbenzene	3.753	3.797	-1.2	106	0.00
79 T	2-Chlorotoluene	2.217	2.209	0.4	105	0.00
80 T	1,3,5-Trimethylbenzene	2.771	2.814	-1.6	105	0.00
81 T	trans-1,4-Dichloro-2-butene	0.245	0.274	-11.8	109	0.00
82 T	4-Chlorotoluene	2.679	2.685	-0.2	106	0.00
83 T	tert-Butylbenzene	2.795	2.877	-2.9	106	0.00
84 T	1,2,4-Trimethylbenzene	2.852	2.888	-1.3	105	0.00
85 T	sec-Butylbenzene	3.337	3.387	-1.5	106	0.00
86 T	p-Isopropyltoluene	3.087	3.151	-2.1	106	0.00
87 T	1,3-Dichlorobenzene	1.776	1.750	1.5	104	0.00
88 T	1,4-Dichlorobenzene	1.823	1.795	1.5	106	0.00
89 T	n-Butylbenzene	2.746	2.802	-2.0	107	0.00

Data Path : W:\HPCHEM1\MSVOA X\Data\VX042318\
Data File : VX001048.D
Acq On : 23 Apr 2018 10:21
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 24 07:11:59 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.437	0.503	-15.1	113	0.00
91 T	1,2-Dichlorobenzene	1.720	1.698	1.3	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.194	0.194	0.0	101	0.00
93 T	1,2,4-Trichlorobenzene	1.444	1.423	1.5	105	0.00
94 T	Hexachlorobutadiene	0.653	0.611	6.4	102	0.00
95 T	Naphthalene	3.550	3.409	4.0	98	0.00
96 T	1,2,3-Trichlorobenzene	1.396	1.361	2.5	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6