

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042418\
 Data File : VX001080.D
 Acq On : 24 Apr 2018 16:46
 Operator : JC/MD
 Sample : J2349-14DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012RW163-180411DL

Quant Time: Apr 25 07:29:44 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	243463	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	335350	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	282554	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	148736	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	170392	55.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.42%	
35) Dibromofluoromethane	5.51	113	131259	51.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.66%	
50) Toluene-d8	8.72	98	491354	51.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.92%	
62) 4-Bromofluorobenzene	11.14	95	161729	40.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.52%	
Target Compounds						
4) Vinyl Chloride	1.40	62	3644	1.250	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	112597	38.278	ug/l	89
65) Chlorobenzene	10.15	112	15625	2.077	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042418\
Data File : VX001080.D
Acq On : 24 Apr 2018 16:46
Operator : JC/MD
Sample : J2349-14DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012RW163-180411DL

Quant Time: Apr 25 07:29:44 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

