

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042418\
 Data File : VX001090.D
 Acq On : 24 Apr 2018 21:53
 Operator : JC/MD
 Sample : J2463-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z1TB07-180416

Quant Time: Apr 25 07:51:27 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	232665	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	324682	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	284857	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	158814	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	168845	57.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.50%	
35) Dibromofluoromethane	5.51	113	129085	52.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.28%	
50) Toluene-d8	8.73	98	482162	52.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.34%	
62) 4-Bromofluorobenzene	11.14	95	163821	42.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.28%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042418\
Data File : VX001090.D
Acq On : 24 Apr 2018 21:53
Operator : JC/MD
Sample : J2463-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z1TB07-180416

Quant Time: Apr 25 07:51:27 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

