

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042418\
 Data File : VX001093.D
 Acq On : 24 Apr 2018 23:09
 Operator : JC/MD
 Sample : J2463-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014RW133-180416

Quant Time: Apr 25 08:00:07 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	227793	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	322832	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	282575	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	157168	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	169530	58.71	ug/l	0.00
Spiked Amount			Recovery	=	117.42%	
35) Dibromofluoromethane	5.51	113	129345	52.54	ug/l	0.00
Spiked Amount			Recovery	=	105.08%	
50) Toluene-d8	8.72	98	490346	53.87	ug/l	0.00
Spiked Amount			Recovery	=	107.74%	
62) 4-Bromofluorobenzene	11.14	95	167159	43.76	ug/l	0.00
Spiked Amount			Recovery	=	87.52%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.41	62	3435	1.26	ug/l	97
21) trans-1,2-Dichloroethene	3.17	96	2999	1.21	ug/l	93
27) cis-1,2-Dichloroethene	4.60	96	46805	17.01	ug/l	89
44) Trichloroethene	7.23	130	7747	2.40	ug/l	87
65) Chlorobenzene	10.15	112	57638	7.66	ug/l	97
87) 1,3-Dichlorobenzene	12.04	146	1597	0.29	ug/l	94
88) 1,4-Dichlorobenzene	12.10	146	15570	2.72	ug/l	91
91) 1,2-Dichlorobenzene	12.40	146	18230	3.37	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042418\
Data File : VX001093.D
Acq On : 24 Apr 2018 23:09
Operator : JC/MD
Sample : J2463-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014RW133-180416

Quant Time: Apr 25 08:00:07 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

