

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042418\
 Data File : VX001097.D
 Acq On : 25 Apr 2018 00:52
 Operator : JC/MD
 Sample : J2463-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS043MW039-180417

Quant Time: Apr 25 08:08:09 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	225186	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	317530	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	279663	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	156195	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	166522	58.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.68%	
35) Dibromofluoromethane	5.51	113	126047	52.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.12%	
50) Toluene-d8	8.72	98	482329	53.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.74%	
62) 4-Bromofluorobenzene	11.14	95	162525	43.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.50%	
Target Compounds						
44) Trichloroethene	7.23	130	1333	0.42	ug/l	Qvalue 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042418\
Data File : VX001097.D
Acq On : 25 Apr 2018 00:52
Operator : JC/MD
Sample : J2463-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS043MW039-180417

Quant Time: Apr 25 08:08:09 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

