

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042418\
 Data File : VX001100.D
 Acq On : 25 Apr 2018 02:08
 Operator : JC/MD
 Sample : J2463-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF015RW009-180416

Quant Time: Apr 25 08:13:46 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	233556	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	321537	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	281652	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	147775	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	167687	56.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.28%	
35) Dibromofluoromethane	5.51	113	128223	52.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.60%	
50) Toluene-d8	8.72	98	491777	54.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.48%	
62) 4-Bromofluorobenzene	11.14	95	161501	42.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.90%	

Target Compounds						Qvalue
16) Acetone	2.45	43	4413	3.96	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042418\
Data File : VX001100.D
Acq On : 25 Apr 2018 02:08
Operator : JC/MD
Sample : J2463-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF015RW009-180416

Quant Time: Apr 25 08:13:46 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

