

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042418\
 Data File : VX001111.D
 Acq On : 25 Apr 2018 06:49
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 25 08:45:03 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.620	0.640	-3.2	95	0.00
3 P	Chloromethane	0.577	0.552	4.3	93	0.00
4 C	Vinyl Chloride	0.598	0.626	-4.7#	104	0.00
5 T	Bromomethane	0.370	0.344	7.0	89	0.00
6 T	Chloroethane	0.356	0.360	-1.1	103	0.00
7 T	Trichlorofluoromethane	0.899	0.927	-3.1	103	0.00
8 T	Diethyl Ether	0.332	0.339	-2.1	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.527	0.516	2.1	100	0.00
10 T	Methyl Iodide	0.437	0.479	-9.6	103	0.00
11 T	Tert butyl alcohol	0.096	0.122	-27.1#	134	0.00
12 CM	1,1-Dichloroethene	0.487	0.483	0.8#	100	0.00
13 T	Acrolein	0.031	0.046	-48.4#	169#	0.00
14 T	Allyl chloride	0.882	0.883	-0.1	101	0.00
15 T	Acrylonitrile	0.259	0.282	-8.9	114	0.00
16 T	Acetone	0.239	0.249	-4.2	110	0.00
17 T	Carbon Disulfide	1.388	1.213	12.6	89	0.00
18 T	Methyl Acetate	0.722	0.917	-27.0#	122	0.00
19 T	Methyl tert-butyl Ether	1.638	1.745	-6.5	108	0.00
20 T	Methylene Chloride	0.556	0.550	1.1	106	0.00
21 T	trans-1,2-Dichloroethene	0.542	0.521	3.9	101	0.00
22 T	Diisopropyl ether	1.707	1.785	-4.6	104	0.00
23 T	Vinyl Acetate	1.413	1.421	-0.6	101	0.00
24 P	1,1-Dichloroethane	0.957	0.944	1.4	101	0.00
25 T	2-Butanone	0.373	0.399	-7.0	112	0.00
26 T	2,2-Dichloropropane	0.823	0.576	30.0#	70	0.00
27 T	cis-1,2-Dichloroethene	0.604	0.589	2.5	99	0.00
28 T	Bromochloromethane	0.368	0.392	-6.5	106	0.00
29 T	Tetrahydrofuran	0.235	0.257	-9.4	113	0.00
30 C	Chloroform	0.991	1.010	-1.9#	102	0.00
31 T	Cyclohexane	0.894	0.834	6.7	97	0.00
32 T	1,1,1-Trichloroethane	0.889	0.917	-3.1	102	0.00
33 S	1,2-Dichloroethane-d4	0.634	0.597	5.8	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.381	0.361	5.2	96	0.00
36 T	1,1-Dichloropropene	0.588	0.538	8.5	97	0.00
37 T	Ethyl Acetate	0.539	0.568	-5.4	110	0.00
38 T	Carbon Tetrachloride	0.595	0.582	2.2	100	0.00
39 T	Methylcyclohexane	0.684	0.620	9.4	94	0.00
40 TM	Benzene	1.634	1.580	3.3	100	0.00
41 T	Methacrylonitrile	0.337	0.338	-0.3	109	0.00
42 TM	1,2-Dichloroethane	0.627	0.625	0.3	105	0.00
43 T	Isopropyl Acetate	0.925	0.961	-3.9	108	0.00
44 TM	Trichloroethene	0.499	0.466	6.6	98	0.00
45 C	1,2-Dichloropropane	0.411	0.411	0.0	101	0.00

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042418\
 Data File : VX001111.D
 Acq On : 25 Apr 2018 06:49
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 25 08:45:03 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.287	0.284	1.0	103	0.00
47 T	Bromodichloromethane	0.532	0.554	-4.1	104	0.00
48 T	Methyl methacrylate	0.498	0.519	-4.2	108	0.00
49 T	1,4-Dioxane	0.007	0.009	-28.6#	140	0.00
50 S	Toluene-d8	1.410	1.266	10.2	92	0.00
51 T	4-Methyl-2-Pentanone	0.561	0.628	-11.9	116	0.00
52 CM	Toluene	1.066	1.055	1.0#	102	0.00
53 T	t-1,3-Dichloropropene	0.636	0.620	2.5	95	0.00
54 T	cis-1,3-Dichloropropene	0.600	0.591	1.5	95	0.00
55 T	1,1,2-Trichloroethane	0.424	0.432	-1.9	105	0.00
56 T	Ethyl methacrylate	0.622	0.653	-5.0	106	0.00
57 T	1,3-Dichloropropane	0.702	0.702	0.0	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.304	0.318	-4.6	105	0.00
59 T	2-Hexanone	0.430	0.476	-10.7	115	0.00
60 T	Dibromochloromethane	0.445	0.474	-6.5	103	0.00
61 T	1,2-Dibromoethane	0.447	0.463	-3.6	105	0.00
62 S	4-Bromofluorobenzene	0.592	0.533	10.0	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.531	0.532	-0.2	106	0.00
65 PM	Chlorobenzene	1.331	1.279	3.9	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.457	0.473	-3.5	102	0.00
67 C	Ethyl Benzene	2.211	2.149	2.8#	101	0.00
68 T	m/p-Xylenes	0.871	0.854	2.0	101	0.00
69 T	o-Xylene	0.840	0.829	1.3	102	0.00
70 T	Styrene	1.384	1.380	0.3	101	0.00
71 P	Bromoform	0.375	0.402	-7.2	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.269	3.255	0.4	103	0.00
74 T	N-amyl acetate	1.398	1.444	-3.3	108	0.00
75 P	1,1,2,2-Tetrachloroethane	0.925	0.997	-7.8	110	0.00
76 T	1,2,3-Trichloropropane	0.867	0.917	-5.8	109	0.00
77 T	Bromobenzene	0.949	0.927	2.3	101	0.00
78 T	n-propylbenzene	3.753	3.658	2.5	101	0.00
79 T	2-Chlorotoluene	2.217	2.194	1.0	103	0.00
80 T	1,3,5-Trimethylbenzene	2.771	2.775	-0.1	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.245	0.230	6.1	90	0.00
82 T	4-Chlorotoluene	2.679	2.627	1.9	103	0.00
83 T	tert-Butylbenzene	2.795	2.829	-1.2	103	0.00
84 T	1,2,4-Trimethylbenzene	2.852	2.867	-0.5	103	0.00
85 T	sec-Butylbenzene	3.337	3.335	0.1	103	0.00
86 T	p-Isopropyltoluene	3.087	3.045	1.4	102	0.00
87 T	1,3-Dichlorobenzene	1.776	1.708	3.8	101	0.00
88 T	1,4-Dichlorobenzene	1.823	1.739	4.6	101	0.00
89 T	n-Butylbenzene	2.746	2.609	5.0	99	0.00

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042418\
Data File : VX001111.D
Acq On : 25 Apr 2018 06:49
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 25 08:45:03 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.437	0.470	-7.6	104	0.00
91 T	1,2-Dichlorobenzene	1.720	1.705	0.9	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.194	0.224	-15.5	115	0.00
93 T	1,2,4-Trichlorobenzene	1.444	1.337	7.4	98	0.00
94 T	Hexachlorobutadiene	0.653	0.649	0.6	107	0.00
95 T	Naphthalene	3.550	3.730	-5.1	106	0.00
96 T	1,2,3-Trichlorobenzene	1.396	1.358	2.7	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5