

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042518\
 Data File : VX001122.D
 Acq On : 25 Apr 2018 16:09
 Operator : JC/MD
 Sample : J2490-02
 Misc : 5.02g/5mL/100uL/5mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MDP-B-3-18-20

Quant Time: Apr 25 19:51:43 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	225282	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	319625	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	292449	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	178300	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	151029	52.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.78%	
35) Dibromofluoromethane	5.51	113	107711	44.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.38%	
50) Toluene-d8	8.72	98	442396	49.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.18%	
62) 4-Bromofluorobenzene	11.15	95	165982	43.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.78%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042518\
Data File : VX001122.D
Acq On : 25 Apr 2018 16:09
Operator : JC/MD
Sample : J2490-02
Misc : 5.02g/5mL/100uL/5mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MDP-B-3-18-20

Quant Time: Apr 25 19:51:43 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

