

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042518\
 Data File : VX001125.D
 Acq On : 25 Apr 2018 17:47
 Operator : JC/MD
 Sample : J2464-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z1TB09-180417

Quant Time: Apr 26 05:12:03 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	231249	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	321534	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	286143	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	161041	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	175262	59.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.58%	
35) Dibromofluoromethane	5.51	113	131950	53.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.64%	
50) Toluene-d8	8.72	98	503669	55.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.12%	
62) 4-Bromofluorobenzene	11.14	95	176539	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
Target Compounds						
4) Vinyl Chloride	1.40	62	699	0.25	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042518\
Data File : VX001125.D
Acq On : 25 Apr 2018 17:47
Operator : JC/MD
Sample : J2464-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z1TB09-180417

Quant Time: Apr 26 05:12:03 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

