

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042518\
 Data File : VX001130.D
 Acq On : 25 Apr 2018 19:55
 Operator : JC/MD
 Sample : J2464-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF017MW004-180417

Quant Time: Apr 26 05:22:24 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	253242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	354315	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	337990	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	191506	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	176070	54.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.70%	
35) Dibromofluoromethane	5.51	113	135267	50.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.14%	
50) Toluene-d8	8.72	98	544898	54.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.08%	
62) 4-Bromofluorobenzene	11.14	95	206571	49.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.54%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042518\
Data File : VX001130.D
Acq On : 25 Apr 2018 19:55
Operator : JC/MD
Sample : J2464-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF017MW004-180417

Quant Time: Apr 26 05:22:24 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

