

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042618\
 Data File : VX001150.D
 Acq On : 27 Apr 2018 03:40
 Operator : JC/MD
 Sample : PB108612ZHE#02 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108612ZHE#02

Quant Time: Apr 27 07:38:58 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	226126	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	314105	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	286714	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	173338	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	146908	41.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.82%	
35) Dibromofluoromethane	5.51	113	114305	43.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.94%	
50) Toluene-d8	8.72	98	445640	45.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.84%	
62) 4-Bromofluorobenzene	11.14	95	158141	41.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.52%	
Target Compounds						
16) Acetone	2.45	43	3633	2.520	ug/l	94
18) Methyl Acetate	2.79	43	3895	1.080	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042618\
Data File : VX001150.D
Acq On : 27 Apr 2018 03:40
Operator : JC/MD
Sample : PB108612ZHE#02 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108612ZHE#02

Quant Time: Apr 27 07:38:58 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

