

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042618\
 Data File : VX001154.D
 Acq On : 27 Apr 2018 05:23
 Operator : JC/MD
 Sample : PB108612ZHE#06 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108612ZHE#06

Quant Time: Apr 27 08:11:38 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	208223	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	289331	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	257595	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	148364	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	136074	42.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.32%	
35) Dibromofluoromethane	5.51	113	105727	44.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.30%	
50) Toluene-d8	8.72	98	403533	44.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.30%	
62) 4-Bromofluorobenzene	11.14	95	139471	39.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.00%	
Target Compounds						
16) Acetone	2.45	43	3664	2.760	ug/l	98
18) Methyl Acetate	2.79	43	3807	1.146	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042618\
Data File : VX001154.D
Acq On : 27 Apr 2018 05:23
Operator : JC/MD
Sample : PB108612ZHE#06 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108612ZHE#06

Quant Time: Apr 27 08:11:38 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

