

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042618\
 Data File : VX001158.D
 Acq On : 27 Apr 2018 07:05
 Operator : JC/MD
 Sample : PB108612ZHE#10 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108612ZHE#10

Quant Time: Apr 27 08:17:48 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	207041	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	286743	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	251809	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133605	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	133921	41.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.46%	
35) Dibromofluoromethane	5.51	113	103036	43.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.82%	
50) Toluene-d8	8.72	98	400025	44.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.32%	
62) 4-Bromofluorobenzene	11.14	95	131701	37.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.28%	
Target Compounds						
16) Acetone	2.45	43	3708	2.810	ug/l	95
18) Methyl Acetate	2.79	43	3951	1.196	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042618\
Data File : VX001158.D
Acq On : 27 Apr 2018 07:05
Operator : JC/MD
Sample : PB108612ZHE#10 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108612ZHE#10

Quant Time: Apr 27 08:17:48 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

