

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
 Data File : VX001172.D
 Acq On : 27 Apr 2018 16:03
 Operator : JC/MD
 Sample : J2463-15DL 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF015MW045-180416DL

Quant Time: Apr 28 03:47:59 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	200868	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	281185	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	251745	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	144003	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	142875	45.89	ug/l	0.00
Spiked Amount			Recovery	=	91.78%	
35) Dibromofluoromethane	5.51	113	111201	47.78	ug/l	0.00
Spiked Amount			Recovery	=	95.56%	
50) Toluene-d8	8.72	98	422534	48.11	ug/l	0.00
Spiked Amount			Recovery	=	96.22%	
62) 4-Bromofluorobenzene	11.14	95	146218	42.62	ug/l	0.00
Spiked Amount			Recovery	=	85.24%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	90284	38.57	ug/l	99
27) cis-1,2-Dichloroethene	4.62	96	4838	1.97	ug/l	90
40) Benzene	6.15	78	11190	1.25	ug/l	94
52) Toluene	8.79	92	2104	0.36	ug/l	96
65) Chlorobenzene	10.15	112	147839	22.41	ug/l	98
87) 1,3-Dichlorobenzene	12.03	146	2603	0.52	ug/l	95
88) 1,4-Dichlorobenzene	12.10	146	19978	3.84	ug/l	95
91) 1,2-Dichlorobenzene	12.40	146	54438	10.78	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
Data File : VX001172.D
Acq On : 27 Apr 2018 16:03
Operator : JC/MD
Sample : J2463-15DL 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF015MW045-180416DL

Quant Time: Apr 28 03:47:59 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

