

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
 Data File : VX001175.D
 Acq On : 27 Apr 2018 17:20
 Operator : JC/MD
 Sample : J2464-07DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0446-180418DL

Quant Time: Apr 28 06:29:04 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	191006	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	260607	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	232706	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	130585	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	134350	45.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.76%	
35) Dibromofluoromethane	5.51	113	105683	48.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.98%	
50) Toluene-d8	8.72	98	405255	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
62) 4-Bromofluorobenzene	11.14	95	137482	43.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.46%	
Target Compounds						
4) Vinyl Chloride	1.40	62	3694	1.66	ug/l	92
40) Benzene	6.15	78	36181	4.36	ug/l	98
65) Chlorobenzene	10.15	112	354132	58.08	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
Data File : VX001175.D
Acq On : 27 Apr 2018 17:20
Operator : JC/MD
Sample : J2464-07DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0446-180418DL

Quant Time: Apr 28 06:29:04 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

