

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
 Data File : VX001181.D
 Acq On : 27 Apr 2018 19:53
 Operator : JC/MD
 Sample : J2463-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0444-180416

Quant Time: Apr 28 04:01:02 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	188796	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	259730	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	226181	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	121607	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	150841	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
35) Dibromofluoromethane	5.51	113	117581	54.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.38%	
50) Toluene-d8	8.72	98	446760	55.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.14%	
62) 4-Bromofluorobenzene	11.14	95	148497	46.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.70%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	3545	2.95	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
Data File : VX001181.D
Acq On : 27 Apr 2018 19:53
Operator : JC/MD
Sample : J2463-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0444-180416

Quant Time: Apr 28 04:01:02 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

