

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
 Data File : VX001183.D
 Acq On : 27 Apr 2018 20:44
 Operator : JC/MD
 Sample : J2464-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS043MW010-180418

Quant Time: Apr 28 05:55:59 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	194764	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	267060	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	241105	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	137232	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	149319	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
35) Dibromofluoromethane	5.52	113	118341	53.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.08%	
50) Toluene-d8	8.72	98	449887	53.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.86%	
62) 4-Bromofluorobenzene	11.14	95	156052	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
Target Compounds						
16) Acetone	2.45	43	4501	3.63	ug/l	97
49) 1,4-Dioxane	7.77	88	871	14.79	ug/l #	81
65) Chlorobenzene	10.15	112	133898	21.20	ug/l	99
88) 1,4-Dichlorobenzene	12.10	146	15826	3.19	ug/l	93
91) 1,2-Dichlorobenzene	12.40	146	2601	0.54	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
Data File : VX001183.D
Acq On : 27 Apr 2018 20:44
Operator : JC/MD
Sample : J2464-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS043MW010-180418

Quant Time: Apr 28 05:55:59 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

