

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
 Data File : VX001193.D
 Acq On : 28 Apr 2018 01:51
 Operator : JC/MD
 Sample : J2464-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z1TB09-180417

Quant Time: Apr 28 04:49:26 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	204956	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	277084	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	247886	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	145008	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	155389	48.91	ug/l	0.00
Spiked Amount			Recovery	=	97.82%	
35) Dibromofluoromethane	5.51	113	122507	53.42	ug/l	0.00
Spiked Amount			Recovery	=	106.84%	
50) Toluene-d8	8.72	98	468174	54.09	ug/l	0.00
Spiked Amount			Recovery	=	108.18%	
62) 4-Bromofluorobenzene	11.14	95	160069	47.34	ug/l	0.00
Spiked Amount			Recovery	=	94.68%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	3024	2.31	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
Data File : VX001193.D
Acq On : 28 Apr 2018 01:51
Operator : JC/MD
Sample : J2464-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z1TB09-180417

Quant Time: Apr 28 04:49:26 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

