

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
 Data File : VX001194.D
 Acq On : 28 Apr 2018 02:16
 Operator : JC/MD
 Sample : J2464-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF017MW001-180417

Quant Time: Apr 28 05:57:23 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	193642	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	266398	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	232596	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	129375	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	151555	50.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.98%	
35) Dibromofluoromethane	5.51	113	120009	54.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.86%	
50) Toluene-d8	8.72	98	451956	54.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.62%	
62) 4-Bromofluorobenzene	11.14	95	149631	46.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.06%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
Data File : VX001194.D
Acq On : 28 Apr 2018 02:16
Operator : JC/MD
Sample : J2464-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF017MW001-180417

Quant Time: Apr 28 05:57:23 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

