

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
 Data File : VX001196.D
 Acq On : 28 Apr 2018 03:07
 Operator : JC/MD
 Sample : J2464-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS043MW100-180418

Quant Time: Apr 28 05:58:23 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	196369	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	267585	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	243949	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	137093	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	150953	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
35) Dibromofluoromethane	5.51	113	119251	53.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.68%	
50) Toluene-d8	8.72	98	452082	54.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.18%	
62) 4-Bromofluorobenzene	11.14	95	158018	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.61	96	1435	0.60	ug/l	89
40) Benzene	6.15	78	219979	25.79	ug/l	98
65) Chlorobenzene	10.15	112	1959763	306.63	ug/l	99
87) 1,3-Dichlorobenzene	12.03	146	1983	0.42	ug/l	99
88) 1,4-Dichlorobenzene	12.10	146	9899	2.00	ug/l	83
91) 1,2-Dichlorobenzene	12.40	146	4442	0.92	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
Data File : VX001196.D
Acq On : 28 Apr 2018 03:07
Operator : JC/MD
Sample : J2464-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS043MW100-180418

Quant Time: Apr 28 05:58:23 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

