

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
 Data File : VX001205.D
 Acq On : 28 Apr 2018 06:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 28 07:18:45 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.524	0.457	12.8	74	0.00
3 P	Chloromethane	0.549	0.514	6.4	80	0.00
4 C	Vinyl Chloride	0.583	0.563	3.4#	80	0.00
5 T	Bromomethane	0.399	0.483	-21.1#	105	0.00
6 T	Chloroethane	0.367	0.367	0.0	81	0.00
7 T	Trichlorofluoromethane	0.941	0.882	6.3	79	0.00
8 T	Diethyl Ether	0.362	0.358	1.1	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.518	0.438	15.4	72	0.00
10 T	Methyl Iodide	0.470	0.408	13.2	66	0.00
11 T	Tert butyl alcohol	0.153	0.119	22.2#	61	-0.01
12 CM	1,1-Dichloroethene	0.506	0.474	6.3#	79	0.00
13 T	Acrolein	0.058	0.040	31.0#	57	0.00
14 T	Allyl chloride	0.922	0.889	3.6	76	0.00
15 T	Acrylonitrile	0.326	0.293	10.1	74	0.00
16 T	Acetone	0.319	0.259	18.8	68	0.00
17 T	Carbon Disulfide	1.290	1.158	10.2	74	0.00
18 T	Methyl Acetate	0.798	0.767	3.9	77	0.00
19 T	Methyl tert-butyl Ether	1.861	1.868	-0.4	82	0.00
20 T	Methylene Chloride	0.594	0.584	1.7	84	0.00
21 T	trans-1,2-Dichloroethene	0.559	0.535	4.3	80	0.00
22 T	Diisopropyl ether	1.846	1.842	0.2	80	0.00
23 T	Vinyl Acetate	1.566	1.544	1.4	76	0.00
24 P	1,1-Dichloroethane	0.970	0.946	2.5	80	0.00
25 T	2-Butanone	0.477	0.397	16.8	68	-0.01
26 T	2,2-Dichloropropane	0.763	0.532	30.3#	57	0.00
27 T	cis-1,2-Dichloroethene	0.612	0.617	-0.8	81	0.00
28 T	Bromochloromethane	0.435	0.463	-6.4	82	0.00
29 T	Tetrahydrofuran	0.307	0.259	15.6	69	0.00
30 C	Chloroform	1.031	1.061	-2.9#	82	0.00
31 T	Cyclohexane	0.824	0.657	20.3#	66	0.00
32 T	1,1,1-Trichloroethane	0.890	0.880	1.1	78	0.00
33 S	1,2-Dichloroethane-d4	0.775	0.744	4.0	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	79	0.00
35 S	Dibromofluoromethane	0.414	0.431	-4.1	84	0.00
36 T	1,1-Dichloropropene	0.551	0.488	11.4	75	0.00
37 T	Ethyl Acetate	0.660	0.575	12.9	72	0.00
38 T	Carbon Tetrachloride	0.565	0.525	7.1	76	0.00
39 T	Methylcyclohexane	0.610	0.451	26.1#	62	0.00
40 TM	Benzene	1.594	1.595	-0.1	81	0.00
41 T	Methacrylonitrile	0.371	0.334	10.0	74	0.00
42 TM	1,2-Dichloroethane	0.642	0.648	-0.9	82	0.00
43 T	Isopropyl Acetate	1.036	0.971	6.3	75	0.00
44 TM	Trichloroethene	0.475	0.463	2.5	80	0.00
45 C	1,2-Dichloropropane	0.419	0.419	0.0	80	0.00

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
 Data File : VX001205.D
 Acq On : 28 Apr 2018 06:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 28 07:18:45 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.290	0.302	-4.1	83	0.00
47 T	Bromodichloromethane	0.527	0.558	-5.9	82	0.00
48 T	Methyl methacrylate	0.551	0.516	6.4	75	0.00
49 T	1,4-Dioxane	0.011	0.009	18.2	62	0.00
50 S	Toluene-d8	1.562	1.515	3.0	80	0.00
51 T	4-Methyl-2-Pentanone	0.681	0.636	6.6	74	0.00
52 CM	Toluene	1.042	1.023	1.8#	79	0.00
53 T	t-1,3-Dichloropropene	0.613	0.621	-1.3	76	0.00
54 T	cis-1,3-Dichloropropene	0.576	0.576	0.0	74	0.00
55 T	1,1,2-Trichloroethane	0.441	0.455	-3.2	84	0.00
56 T	Ethyl methacrylate	0.653	0.653	0.0	78	0.00
57 T	1,3-Dichloropropane	0.713	0.732	-2.7	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.309	0.304	1.6	78	0.00
59 T	2-Hexanone	0.533	0.478	10.3	71	0.00
60 T	Dibromochloromethane	0.445	0.489	-9.9	82	0.00
61 T	1,2-Dibromoethane	0.462	0.480	-3.9	81	0.00
62 S	4-Bromofluorobenzene	0.610	0.597	2.1	79	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.534	0.521	2.4	83	0.00
65 PM	Chlorobenzene	1.310	1.298	0.9	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.453	0.480	-6.0	81	0.00
67 C	Ethyl Benzene	2.128	2.013	5.4#	77	0.00
68 T	m/p-Xylenes	0.841	0.804	4.4	77	0.00
69 T	o-Xylene	0.816	0.800	2.0	78	0.00
70 T	Styrene	1.358	1.368	-0.7	80	0.00
71 P	Bromoform	0.395	0.415	-5.1	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.259	2.966	9.0	75	0.00
74 T	N-amyl acetate	1.642	1.504	8.4	75	0.00
75 P	1,1,2,2-Tetrachloroethane	1.081	1.054	2.5	80	0.00
76 T	1,2,3-Trichloropropane	1.018	0.976	4.1	77	0.00
77 T	Bromobenzene	0.966	0.959	0.7	82	0.00
78 T	n-propylbenzene	3.634	3.315	8.8	74	0.00
79 T	2-Chlorotoluene	2.253	2.094	7.1	77	0.00
80 T	1,3,5-Trimethylbenzene	2.759	2.597	5.9	76	0.00
81 T	trans-1,4-Dichloro-2-butene	0.256	0.224	12.5	64	0.00
82 T	4-Chlorotoluene	2.678	2.499	6.7	76	0.00
83 T	tert-Butylbenzene	2.754	2.455	10.9	71	0.00
84 T	1,2,4-Trimethylbenzene	2.836	2.714	4.3	76	0.00
85 T	sec-Butylbenzene	3.217	2.839	11.8	72	0.00
86 T	p-Isopropyltoluene	2.992	2.663	11.0	72	0.00
87 T	1,3-Dichlorobenzene	1.735	1.680	3.2	80	0.00
88 T	1,4-Dichlorobenzene	1.808	1.726	4.5	80	0.00
89 T	n-Butylbenzene	2.590	2.192	15.4	69	0.00

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042718\
Data File : VX001205.D
Acq On : 28 Apr 2018 06:57
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 28 07:18:45 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.424	0.393	7.3	72	0.00
91 T	1,2-Dichlorobenzene	1.754	1.742	0.7	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.239	0.225	5.9	72	0.00
93 T	1,2,4-Trichlorobenzene	1.375	1.277	7.1	77	0.00
94 T	Hexachlorobutadiene	0.646	0.507	21.5#	66	0.00
95 T	Naphthalene	3.830	3.760	1.8	78	0.00
96 T	1,2,3-Trichlorobenzene	1.368	1.315	3.9	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5