

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
 Data File : VX001214.D
 Acq On : 28 Apr 2018 15:30
 Operator : JC/MD
 Sample : J2600-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB01-20180424

Quant Time: Apr 30 02:06:42 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	184128	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	248417	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	223754	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	128544	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	140088	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
35) Dibromofluoromethane	5.51	113	111341	54.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.30%	
50) Toluene-d8	8.72	98	418043	53.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.74%	
62) 4-Bromofluorobenzene	11.14	95	142199	46.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.82%	
Target Compounds						
16) Acetone	2.45	43	7491	6.38	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
Data File : VX001214.D
Acq On : 28 Apr 2018 15:30
Operator : JC/MD
Sample : J2600-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB01-20180424

Quant Time: Apr 30 02:06:42 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

