

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
 Data File : VX001233.D
 Acq On : 29 Apr 2018 00:28
 Operator : JC/MD
 Sample : J2600-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-AOC22-MW03-20180424

Quant Time: Apr 30 04:02:34 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	175952	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	240167	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	214154	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	124678	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	131462	48.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.40%	
35) Dibromofluoromethane	5.51	113	104258	52.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.90%	
50) Toluene-d8	8.72	98	389893	51.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.94%	
62) 4-Bromofluorobenzene	11.14	95	132740	45.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.58%	
Target Compounds						
16) Acetone	2.45	43	33425	29.80	ug/l	97
25) 2-Butanone	4.73	43	9393	5.60	ug/l	96
27) cis-1,2-Dichloroethene	4.60	96	2085	0.97	ug/l	86
44) Trichloroethene	7.22	130	3146	1.38	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
Data File : VX001233.D
Acq On : 29 Apr 2018 00:28
Operator : JC/MD
Sample : J2600-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-AOC22-MW03-20180424

Quant Time: Apr 30 04:02:34 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

