

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
 Data File : VX001236.D
 Acq On : 29 Apr 2018 01:45
 Operator : JC/MD
 Sample : J2600-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-AOC22-MW07-20180424

Quant Time: Apr 30 03:34:02 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	174268	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	240191	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	220117	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	130314	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	133143	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
35) Dibromofluoromethane	5.51	113	106243	53.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.88%	
50) Toluene-d8	8.72	98	404023	53.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.70%	
62) 4-Bromofluorobenzene	11.14	95	139667	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
Target Compounds						
16) Acetone	2.45	43	4621	4.16	ug/l	Qvalue 100
44) Trichloroethene	7.23	130	957	0.42	ug/l	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
Data File : VX001236.D
Acq On : 29 Apr 2018 01:45
Operator : JC/MD
Sample : J2600-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-AOC22-MW07-20180424

Quant Time: Apr 30 03:34:02 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

