

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
 Data File : VX001241.D
 Acq On : 29 Apr 2018 03:53
 Operator : JC/MD
 Sample : J2600-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-DUP01-20180424

Quant Time: Apr 30 03:41:59 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	186945	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	257097	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	228418	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	132449	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	141691	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
35) Dibromofluoromethane	5.51	113	113621	53.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.78%	
50) Toluene-d8	8.72	98	430138	53.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.12%	
62) 4-Bromofluorobenzene	11.14	95	144654	46.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.22%	
Target Compounds						
16) Acetone	2.45	43	24946	20.93	ug/l	98
25) 2-Butanone	4.72	43	7959	4.46	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	1687	0.74	ug/l	85
44) Trichloroethene	7.23	130	3111	1.27	ug/l	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
Data File : VX001241.D
Acq On : 29 Apr 2018 03:53
Operator : JC/MD
Sample : J2600-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-DUP01-20180424

Quant Time: Apr 30 03:41:59 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

