

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
 Data File : VX001247.D
 Acq On : 29 Apr 2018 06:27
 Operator : JC/MD
 Sample : J2600-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-DUP02-20180425

Quant Time: Apr 30 03:51:47 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	181366	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	247551	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	220829	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	117368	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	137580	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
35) Dibromofluoromethane	5.51	113	109611	53.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.00%	
50) Toluene-d8	8.72	98	417392	53.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.96%	
62) 4-Bromofluorobenzene	11.14	95	135327	44.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.60%	
Target Compounds						
16) Acetone	2.45	43	5598	4.84	ug/l	96
32) 1,1,1-Trichloroethane	5.51	97	2218	0.69	ug/l #	50
44) Trichloroethene	7.22	130	15238	6.48	ug/l	94
64) Tetrachloroethene	9.34	164	1429	0.61	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
Data File : VX001247.D
Acq On : 29 Apr 2018 06:27
Operator : JC/MD
Sample : J2600-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-DUP02-20180425

Quant Time: Apr 30 03:51:47 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

