

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
 Data File : VX001249.D
 Acq On : 29 Apr 2018 07:17
 Operator : JC/MD
 Sample : J2600-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-FW-MW02-20180425

Quant Time: Apr 30 03:54:44 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	176711	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	237345	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	214339	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	118456	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	132669	48.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.86%	
35) Dibromofluoromethane	5.51	113	105987	53.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.90%	
50) Toluene-d8	8.72	98	399849	53.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.86%	
62) 4-Bromofluorobenzene	11.14	95	130177	44.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.90%	
Target Compounds						
16) Acetone	2.45	43	5935	5.27	ug/l	89
24) 1,1-Dichloroethane	3.71	63	2670	0.78	ug/l #	95
27) cis-1,2-Dichloroethene	4.60	96	933	0.43	ug/l	80
32) 1,1,1-Trichloroethane	5.51	97	3677	1.17	ug/l #	50
44) Trichloroethene	7.22	130	8457	3.75	ug/l	94
64) Tetrachloroethene	9.34	164	3118	1.36	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
Data File : VX001249.D
Acq On : 29 Apr 2018 07:17
Operator : JC/MD
Sample : J2600-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-FW-MW02-20180425

Quant Time: Apr 30 03:54:44 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

