

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
 Data File : VX001250.D
 Acq On : 29 Apr 2018 07:43
 Operator : JC/MD
 Sample : J2600-21
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-FW-MW03-20180425

Quant Time: Apr 30 03:56:39 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 27 01:51:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	182621	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	248964	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	224602	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	125540	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	139319	49.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.42%	
35) Dibromofluoromethane	5.51	113	110696	53.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.44%	
50) Toluene-d8	8.72	98	415578	53.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.88%	
62) 4-Bromofluorobenzene	11.14	95	135069	44.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.92%	
Target Compounds						
16) Acetone	2.45	43	6828	5.87	ug/l	99
44) Trichloroethene	7.22	130	843	0.36	ug/l	85
64) Tetrachloroethene	9.34	164	932	0.39	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042818\
Data File : VX001250.D
Acq On : 29 Apr 2018 07:43
Operator : JC/MD
Sample : J2600-21
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-FW-MW03-20180425

Quant Time: Apr 30 03:56:39 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X042618W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 27 01:51:42 2018
Response via : Initial Calibration

