


Data Path : W:\HPCHEM1\MSVOA X\Data\VX042818\  
Data File : VX001206.D  
Acq On : 28 Apr 2018 10:42  
Operator : JC/MD  
Sample : BFB  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA\_X\METHOD\82X042618W.M  
Title : SW846 8260  
Last Update : Fri Apr 27 01:51:42 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1724

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 20.6      | 3729    | PASS             |
| 75          | 95           | 30           | 60           | 49.1      | 8877    | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 18069   | PASS             |
| 96          | 95           | 5            | 9            | 6.9       | 1254    | PASS             |
| 173         | 174          | 0.00         | 2            | 1.3       | 242     | PASS             |
| 174         | 95           | 50           | 100          | 99.4      | 17963   | PASS             |
| 175         | 174          | 5            | 9            | 7.7       | 1387    | PASS             |
| 176         | 174          | 95           | 101          | 95.3      | 17113   | PASS             |
| 177         | 176          | 5            | 9            | 6.6       | 1138    | PASS             |