

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050118\
 Data File : VX001290.D
 Acq On : 01 May 2018 21:39
 Operator : JC/MD
 Sample : J2507-06DL 1000X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS043MW009-180419DL

Quant Time: May 02 05:40:50 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	169585	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	220779	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	195173	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	109821	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	124584	53.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.14%	
35) Dibromofluoromethane	5.51	113	102579	53.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.86%	
50) Toluene-d8	8.72	98	374954	55.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.36%	
62) 4-Bromofluorobenzene	11.14	95	123590	45.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.96%	
Target Compounds						
4) Vinyl Chloride	1.40	62	17201	8.87	ug/l	96
21) trans-1,2-Dichloroethene	3.17	96	2124	1.15	ug/l	89
27) cis-1,2-Dichloroethene	4.60	96	183106	89.98	ug/l	89
91) 1,2-Dichlorobenzene	12.40	146	9813	2.49	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050118\
Data File : VX001290.D
Acq On : 01 May 2018 21:39
Operator : JC/MD
Sample : J2507-06DL 1000X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS043MW009-180419DL

Quant Time: May 02 05:40:50 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

