

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050118\
 Data File : VX001296.D
 Acq On : 02 May 2018 00:20
 Operator : JC/MD
 Sample : PB108752TBZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108752TBZHE#04

Quant Time: May 02 06:38:14 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	170510	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	230896	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	208561	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	118030	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	104787	44.40	ug/l	0.00
Spiked Amount			Recovery	=	88.80%	
35) Dibromofluoromethane	5.51	113	86691	43.58	ug/l	0.00
Spiked Amount			Recovery	=	87.16%	
50) Toluene-d8	8.72	98	324613	45.68	ug/l	0.00
Spiked Amount			Recovery	=	91.36%	
62) 4-Bromofluorobenzene	11.14	95	106136	37.34	ug/l	0.00
Spiked Amount			Recovery	=	74.68%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	8344	8.28	ug/l	100
18) Methyl Acetate	2.79	43	15801	6.64	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050118\
Data File : VX001296.D
Acq On : 02 May 2018 00:20
Operator : JC/MD
Sample : PB108752TBZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108752TBZHE#04

Quant Time: May 02 06:38:14 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

