

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050118\
 Data File : VX001301.D
 Acq On : 02 May 2018 02:33
 Operator : JC/MD
 Sample : PB108752TBZHE#10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108752TBZHE#10

Quant Time: May 02 06:45:43 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	166383	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	225480	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	200013	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	109483	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	103193	44.80	ug/l	0.00
Spiked Amount 50.000			Recovery =	89.60%		
35) Dibromofluoromethane	5.51	113	86523	44.54	ug/l	0.00
Spiked Amount 50.000			Recovery =	89.08%		
50) Toluene-d8	8.72	98	316879	45.66	ug/l	0.00
Spiked Amount 50.000			Recovery =	91.32%		
62) 4-Bromofluorobenzene	11.14	95	101973	36.74	ug/l	0.00
Spiked Amount 50.000			Recovery =	73.48%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	8369	8.51	ug/l	98
18) Methyl Acetate	2.78	43	16661	7.18	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050118\
Data File : VX001301.D
Acq On : 02 May 2018 02:33
Operator : JC/MD
Sample : PB108752TBZHE#10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108752TBZHE#10

Quant Time: May 02 06:45:43 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

