

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050118\
 Data File : VX001301.D
 Acq On : 02 May 2018 02:33
 Operator : JC/MD
 Sample : PB108752TBZHE#10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108752TBZHE#10

Quant Time: May 02 06:45:43 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	166383	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	225480	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	200013	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	109483	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	103193	44.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.60%	
35) Dibromofluoromethane	5.51	113	86523	44.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.08%	
50) Toluene-d8	8.72	98	316879	45.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.32%	
62) 4-Bromofluorobenzene	11.14	95	101973	36.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.48%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	8369	8.51	ug/l	98
18) Methyl Acetate	2.78	43	16661	7.18	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050118\
Data File : VX001301.D
Acq On : 02 May 2018 02:33
Operator : JC/MD
Sample : PB108752TBZHE#10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108752TBZHE#10

Quant Time: May 02 06:45:43 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 834

Delta R.T. 0.00 min

Lab File: VX001301.D

Acq: 02 May 2018 02:33

Instrument :

MSVOA_X

ClientSampleId :


PB108752TBZHE#10

Tot Ion:168 Resp: 166383

Ion Ratio Lower Upper




168 100

99 48.2 36.5 54.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#16

Acetone

Concen: 8.51 ug/l

RT: 2.45 min Scan# 306

Delta R.T. 0.01 min

Lab File: VX001301.D


Acq: 02 May 2018 02:33

Tgt Ion: 43 Resp: 8369

Ion Ratio Lower Upper



43 100


58 29.6 22.8 34.2



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#18
Methvl Acetate
Concen: 7.18 ug/l
RT: 2.78 min Scan# 360
Delta R.T. 0.00 min
Lab File: VX001301.D
Acq: 02 May 2018 02:33




Instrument :
MSVOA_X
ClientSampleId :
PB108752TBZHE#10


Tot Ion: 43 Resp: 16661
Ion Ratio Lower Upper
43 100
74 23.1 17.8 26.6



#33
1,2-Dichloroethane-d4
Concen: 44.80 ug/l
RT: 6.07 min Scan# 900
Delta R.T. 0.00 min
Lab File: VX001301.D
Acq: 02 May 2018 02:33

Tgt Ion: 65 Resp: 103193
Ion Ratio Lower Upper
65 100
67 52.1 0.0 102.8





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. 0.00 min

Lab File: VX001301.D

Acq: 02 May 2018 02:33

Instrument :

MSVOA_X

ClientSampleId :

PB108752TBZHE#10


Tot Ion:114 Resp: 225480

Ion Ratio Lower Upper

114 100

63 19.3 0.0 38.8




88 15.1 0.0 32.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 44.54 ug/l

RT: 5.51 min Scan# 808

Delta R.T. 0.00 min

Lab File: VX001301.D

Acq: 02 May 2018 02:33


Tgt Ion:113 Resp: 86523

Ion Ratio Lower Upper

113 100

111 101.3 81.2 121.8



192 25.6 20.7 31.1




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 45.66 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. 0.00 min

Lab File: VX001301.D

Acq: 02 May 2018 02:33

Instrument :

MSVOA_X

ClientSampleId :


PB108752TBZHE#10

Tot Ion: 98 Resp: 316879

Ion Ratio Lower Upper




98 100

100 63.4 51.0 76.4



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V



#62

4-Bromofluorobenzene

Concen: 36.74 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. 0.00 min

Lab File: VX001301.D

Acq: 02 May 2018 02:33


Tgt Ion: 95 Resp: 101973

Ion Ratio Lower Upper

95 100

174 104.2 0.0 201.8



176 99.9 0.0 197.2




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V





#63

Chlorobenzene-d5

Concen: 50.00 ug/l

RT: 10.12 min Scan# 1563

Delta R.T. 0.00 min

Lab File: VX001301.D

Acq: 02 May 2018 02:33

Instrument :

MSVOA_X

ClientSampleId :

PB108752TBZHE#10


Tot Ion:117 Resp: 200013

Ion Ratio Lower Upper




117 100

82 51.9 41.4 62.0

119 32.0 25.5 38.3



Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.09 min Scan# 1886

Delta R.T. 0.00 min

Lab File: VX001301.D

Acq: 02 May 2018 02:33


Tgt Ion:152 Resp: 109483

Ion Ratio Lower Upper

152 100

115 52.1 38.6 115.7

150 155.9 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

