


Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050118\
Data File : VX001283.D
Acq On : 01 May 2018 17:17
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Title : SW846 8260
Last Update : Wed May 02 01:48:25 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1724

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	2998	PASS
75	95	30	60	52.2	7583	PASS
95	95	100	100	100.0	14515	PASS
96	95	5	9	7.5	1084	PASS
173	174	0.00	2	1.7	237	PASS
174	95	50	100	98.0	14222	PASS
175	174	5	9	8.3	1176	PASS
176	174	95	101	96.7	13756	PASS
177	176	5	9	7.5	1032	PASS