

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050218\
 Data File : VX001318.D
 Acq On : 02 May 2018 15:46
 Operator : JC/MD
 Sample : J2600-22
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-HN-MW271-20180425

Quant Time: May 03 05:23:00 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	158555	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	212971	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	190743	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	106406	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	118716	54.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.18%	
35) Dibromofluoromethane	5.51	113	97355	53.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.12%	
50) Toluene-d8	8.72	98	358050	54.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.24%	
62) 4-Bromofluorobenzene	11.14	95	116415	44.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	3142	3.35	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050218\
Data File : VX001318.D
Acq On : 02 May 2018 15:46
Operator : JC/MD
Sample : J2600-22
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-HN-MW271-20180425

Quant Time: May 03 05:23:00 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

