

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050218\
 Data File : VX001321.D
 Acq On : 02 May 2018 17:07
 Operator : JC/MD
 Sample : J2628-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGEBLANK-WATERREF-4

Quant Time: May 03 05:33:23 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	151416	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	201585	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	172341	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	84055	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	93489	44.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.20%	
35) Dibromofluoromethane	5.51	113	75906	43.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.42%	
50) Toluene-d8	8.72	98	276441	44.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.10%	
62) 4-Bromofluorobenzene	11.14	95	83248	33.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.10%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	7924	8.86	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050218\
Data File : VX001321.D
Acq On : 02 May 2018 17:07
Operator : JC/MD
Sample : J2628-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGEBLANK-WATERREF-4

Quant Time: May 03 05:33:23 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

