

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050218\
 Data File : VX001338.D
 Acq On : 03 May 2018 01:38
 Operator : JC/MD
 Sample : PB108794ZHE#17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108794ZHE#17

Quant Time: May 03 06:46:53 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	149104	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	199960	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	175553	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	94839	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	93757	45.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.86%	
35) Dibromofluoromethane	5.51	113	76366	44.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.66%	
50) Toluene-d8	8.72	98	280923	45.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.30%	
62) 4-Bromofluorobenzene	11.15	95	87864	35.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.40%	
Target Compounds						
16) Acetone	2.45	43	11670	13.25	ug/l	97
18) Methyl Acetate	2.79	43	16930	8.14	ug/l	99
20) Methylene Chloride	2.86	84	4526	1.30	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050218\
Data File : VX001338.D
Acq On : 03 May 2018 01:38
Operator : JC/MD
Sample : PB108794ZHE#17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108794ZHE#17

Quant Time: May 03 06:46:53 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

