

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050218\
 Data File : VX001340.D
 Acq On : 03 May 2018 02:32
 Operator : JC/MD
 Sample : PB108794ZHE#19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108794ZHE#19

Quant Time: May 03 06:49:39 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	154348	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	206874	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	188702	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	106969	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	95304	44.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.22%	
35) Dibromofluoromethane	5.51	113	79449	44.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.16%	
50) Toluene-d8	8.72	98	289054	45.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.80%	
62) 4-Bromofluorobenzene	11.14	95	94819	37.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.48%	
Target Compounds						
16) Acetone	2.45	43	12227	13.41	ug/l	97
18) Methyl Acetate	2.79	43	17256	8.01	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050218\
Data File : VX001340.D
Acq On : 03 May 2018 02:32
Operator : JC/MD
Sample : PB108794ZHE#19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108794ZHE#19

Quant Time: May 03 06:49:39 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

