

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050218\
 Data File : VX001343.D
 Acq On : 03 May 2018 03:53
 Operator : JC/MD
 Sample : PB108794ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108794ZHE#02

Quant Time: May 03 06:52:19 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	148264	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	201623	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	178950	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	95665	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	92961	45.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.58%	
35) Dibromofluoromethane	5.51	113	77057	44.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.72%	
50) Toluene-d8	8.72	98	284290	45.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.62%	
62) 4-Bromofluorobenzene	11.14	95	88377	35.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.22%	
Target Compounds						
16) Acetone	2.45	43	12285	14.02	ug/l	96
18) Methyl Acetate	2.79	43	17466	8.44	ug/l	98
20) Methylene Chloride	2.86	84	4169	1.08	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050218\
Data File : VX001343.D
Acq On : 03 May 2018 03:53
Operator : JC/MD
Sample : PB108794ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108794ZHE#02

Quant Time: May 03 06:52:19 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

