

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050218\
 Data File : VX001347.D
 Acq On : 03 May 2018 05:40
 Operator : JC/MD
 Sample : PB108794ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108794ZHE#06

Quant Time: May 03 06:57:00 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	151371	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	207584	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	190161	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	104906	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	95919	45.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.56%	
35) Dibromofluoromethane	5.51	113	78790	44.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.12%	
50) Toluene-d8	8.72	98	293812	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
62) 4-Bromofluorobenzene	11.14	95	95828	37.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.00%	
Target Compounds						
16) Acetone	2.45	43	11527	12.89	ug/l	93
18) Methyl Acetate	2.79	43	17270	8.18	ug/l	98
20) Methylene Chloride	2.86	84	4167	1.02	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050218\
Data File : VX001347.D
Acq On : 03 May 2018 05:40
Operator : JC/MD
Sample : PB108794ZHE#06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108794ZHE#06

Quant Time: May 03 06:57:00 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

