


Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050218\
Data File : VX001308.D
Acq On : 02 May 2018 09:21
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\624X043018W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Tue May 01 09:17:40 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1725

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	2460	PASS
75	95	30	60	49.2	6573	PASS
95	95	100	100	100.0	13371	PASS
96	95	5	9	6.1	809	PASS
173	174	0.00	2	1.3	169	PASS
174	95	50	100	98.7	13200	PASS
175	174	5	9	7.9	1045	PASS
176	174	95	101	96.1	12688	PASS
177	176	5	9	7.1	895	PASS