

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050318\  
 Data File : VX001367.D  
 Acq On : 03 May 2018 18:50  
 Operator : JC/MD  
 Sample : J2712-03 5X  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampled :  
 0011

Quant Time: May 04 08:07:26 2018  
 Quant Method : W:\HPCHEM1\MSVOA\_X\METHOD\82X050118W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 02 01:48:25 2018  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	151992	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	207779	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	183754	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	83616	50.00	ug/l	0.00

## System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.08	65	79523	37.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	75.60%	
35) Dibromofluoromethane	5.51	113	66962	37.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	74.82%	
50) Toluene-d8	8.73	98	243472	38.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	76.14%	
62) 4-Bromofluorobenzene	11.15	95	74118	28.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	57.96%	

## Target Compounds

						Qvalue
16) Acetone	2.45	43	3671	4.09	ug/l	90
18) Methyl Acetate	2.79	43	4233	2.00	ug/l	98
44) Trichloroethene	7.22	130	4436	2.10	ug/l	88
68) m/p-Xylenes	10.37	106	3910	1.21	ug/l	93
84) 1,2,4-Trimethylbenzene	11.82	105	9516	1.98	ug/l	100
93) 1,2,4-Trichlorobenzene	13.65	180	2445	1.01	ug/l	97
95) Naphthalene	13.84	128	201247	32.89	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050318\  
 Data File : VX001367.D  
 Acq On : 03 May 2018 18:50  
 Operator : JC/MD  
 Sample : J2712-03 5X  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 0011

Quant Time: May 04 08:07:26 2018  
 Quant Method : W:\HPCHEM1\MSVOA\_X\METHOD\82X050118W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 02 01:48:25 2018  
 Response via : Initial Calibration

