

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050318\
 Data File : VX001368.D
 Acq On : 03 May 2018 19:17
 Operator : JC/MD
 Sample : J2713-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1

Quant Time: May 04 08:09:31 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	152713	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	215351	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	194128	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	97570	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.08	65	80155	37.92	ug/l	0.00
Spiked Amount	50.000		Recovery	= 75.84%		
35) Dibromofluoromethane	5.51	113	67164	36.20	ug/l	0.00
Spiked Amount	50.000		Recovery	= 72.40%		
50) Toluene-d8	8.73	98	252600	38.11	ug/l	0.00
Spiked Amount	50.000		Recovery	= 76.22%		
62) 4-Bromofluorobenzene	11.15	95	84061	31.71	ug/l	0.00
Spiked Amount	50.000		Recovery	= 63.42%		

Target Compounds

						Qvalue
16) Acetone	2.45	43	4489	4.97	ug/l	94
18) Methyl Acetate	2.79	43	3666	1.72	ug/l	93
20) Methylene Chloride	2.86	84	40361	23.96	ug/l	99
32) 1,1,1-Trichloroethane	5.50	97	5723	2.15	ug/l	94
44) Trichloroethene	7.23	130	101255	46.26	ug/l	97
67) Ethyl Benzene	10.26	91	12438	1.43	ug/l	92
68) m/p-Xylenes	10.37	106	23164	6.77	ug/l	99
69) o-Xylene	10.71	106	9712	2.93	ug/l	96
95) Naphthalene	13.84	128	35426	4.96	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050318\
 Data File : VX001368.D
 Acq On : 03 May 2018 19:17
 Operator : JC/MD
 Sample : J2713-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 Client Sampled :
 TP-1

Quant Time: May 04 08:09:31 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

