

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050318\
 Data File : VX001372.D
 Acq On : 03 May 2018 21:06
 Operator : JC/MD
 Sample : J2670-02 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-04-050218

Quant Time: May 04 08:17:21 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	152889	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	213476	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	193590	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	97685	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	80022	37.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.62%	
35) Dibromofluoromethane	5.52	113	66824	36.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.66%	
50) Toluene-d8	8.72	98	251081	38.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.42%	
62) 4-Bromofluorobenzene	11.15	95	81643	31.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	62.14%	
Target Compounds						
16) Acetone	2.45	43	4098	4.54	ug/l	93
18) Methyl Acetate	2.79	43	3827	1.79	ug/l	97
44) Trichloroethene	7.23	130	12348	5.69	ug/l	97
68) m/p-Xylenes	10.37	106	6153	1.80	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050318\
Data File : VX001372.D
Acq On : 03 May 2018 21:06
Operator : JC/MD
Sample : J2670-02 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-04-050218

Quant Time: May 04 08:17:21 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

