

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050318\
 Data File : VX001373.D
 Acq On : 03 May 2018 21:33
 Operator : JC/MD
 Sample : J2679-02 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EO-01-050218-A

Quant Time: May 04 08:19:15 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	153457	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	212008	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	191468	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	94850	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	79707	37.52	ug/l	0.00
Spiked Amount			Recovery	=	75.04%	
35) Dibromofluoromethane	5.52	113	66866	36.61	ug/l	0.00
Spiked Amount			Recovery	=	73.22%	
50) Toluene-d8	8.72	98	248867	38.14	ug/l	0.00
Spiked Amount			Recovery	=	76.28%	
62) 4-Bromofluorobenzene	11.15	95	81653	31.29	ug/l	0.00
Spiked Amount			Recovery	=	62.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	3243	3.58	ug/l	94
18) Methyl Acetate	2.79	43	4109	1.92	ug/l	96
20) Methylene Chloride	2.86	84	21191	11.73	ug/l	98
44) Trichloroethene	7.23	130	10170	4.72	ug/l	94
67) Ethyl Benzene	10.26	91	14447	1.68	ug/l	98
68) m/p-Xylenes	10.36	106	27351	8.10	ug/l	98
69) o-Xylene	10.71	106	10507	3.21	ug/l	96
95) Naphthalene	13.84	128	9708	1.40	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050318\
Data File : VX001373.D
Acq On : 03 May 2018 21:33
Operator : JC/MD
Sample : J2679-02 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EO-01-050218-A

Quant Time: May 04 08:19:15 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

