


Data Path : W:\HPCHEM1\MSVOA X\Data\VX050318\
Data File : VX001352.D
Acq On : 03 May 2018 10:57
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Title : SW846 8260
Last Update : Wed May 02 01:48:25 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1724

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.4	2439	PASS
75	95	30	60	48.3	5789	PASS
95	95	100	100	100.0	11980	PASS
96	95	5	9	6.2	744	PASS
173	174	0.00	2	1.6	184	PASS
174	95	50	100	98.9	11851	PASS
175	174	5	9	7.8	928	PASS
176	174	95	101	97.5	11558	PASS
177	176	5	9	6.7	779	PASS