

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050418\
 Data File : VX001395.D
 Acq On : 04 May 2018 13:55
 Operator : JC/MD
 Sample : VX0504WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0504WBL02

Quant Time: May 05 05:54:23 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	163505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	218309	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	194946	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	110826	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	117570	51.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.90%	
35) Dibromofluoromethane	5.51	113	98272	52.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.50%	
50) Toluene-d8	8.72	98	355463	52.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.80%	
62) 4-Bromofluorobenzene	11.14	95	115556	43.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.00%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050418\
Data File : VX001395.D
Acq On : 04 May 2018 13:55
Operator : JC/MD
Sample : VX0504WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0504WBL02

Quant Time: May 05 05:54:23 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

