

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050418\
 Data File : VX001398.D
 Acq On : 04 May 2018 15:50
 Operator : JC/MD
 Sample : J2707-10
 Misc : 6.04g/5mL/100uL/5mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-8(6.5-7)

Quant Time: May 05 05:58:58 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	180486	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	234102	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	208983	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	118436	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	116317	46.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.12%	
35) Dibromofluoromethane	5.51	113	101496	50.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.64%	
50) Toluene-d8	8.72	98	343821	47.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.44%	
62) 4-Bromofluorobenzene	11.14	95	113098	39.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	78.50%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050418\
Data File : VX001398.D
Acq On : 04 May 2018 15:50
Operator : JC/MD
Sample : J2707-10
Misc : 6.04g/5mL/100uL/5mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-8(6.5-7)

Quant Time: May 05 05:58:58 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

