

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050418\
 Data File : VX001400.D
 Acq On : 04 May 2018 16:44
 Operator : JC/MD
 Sample : J2707-09
 Misc : 6.86g/5mL/100uL/5mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-13(5.5-6)

Quant Time: May 05 06:02:08 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	160710	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	209253	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	184382	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	96926	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	111371	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
35) Dibromofluoromethane	5.51	113	91191	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
50) Toluene-d8	8.72	98	315225	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
62) 4-Bromofluorobenzene	11.15	95	97410	37.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.64%	
Target Compounds						
64) Tetrachloroethene	9.34	164	2255	1.032	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050418\
Data File : VX001400.D
Acq On : 04 May 2018 16:44
Operator : JC/MD
Sample : J2707-09
Misc : 6.86g/5mL/100uL/5mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-13(5.5-6)

Quant Time: May 05 06:02:08 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

